



MISMOUR DEPARTMENT OF HEALTH AND SENIOR SERVICES  
STATE PUBLIC HEALTH LABORATORY  
BREATH ALCOHOL PROGRAM  
DATAMASTER MAINTENANCE REPORT

RECEIVED

By Carol Day at 8:31 am, Jul 08, 2013

REPORT #

Complete this report at the time of the regular monthly preventive maintenance check (not to exceed 30 days).  
Complete this report whenever the instrument is serviced or repaired and whenever it is placed into service.  
Retain the original and send a copy within 15 days to the Breath Alcohol Program, DHSS.

DATAMASTER SN 204136	NAME OF AGENCY Missouri State Highway Patrol	DATE OF INSPECTION 07/03/2013
LOCATION OF INSTRUMENT (STREET AND CITY) PIKE COUNTY SHERIFF'S DEPT., BOWLING GREEN, MO, 63334		TIME OF INSPECTION 6:37 am

**CHECKLIST:** Place a mark in the box by each item if found to be satisfactory or if operating within established limits. (Write in observed values where determined.) Unmarked items must be corrected before using instrument.

<input checked="" type="checkbox"/> DIAGNOSTIC CHECK (PRINTOUT ATTACHED)	DATE AND TIME (from printout) 07/03/2013 @ 06:40 AM	
<input checked="" type="checkbox"/> COMPUTER	<input checked="" type="checkbox"/> DETECTOR	
<input checked="" type="checkbox"/> PROGRAM	<input checked="" type="checkbox"/> FILTERS	
<input checked="" type="checkbox"/> HEATERS SAMPLE CHAMBER 49°C	<input checked="" type="checkbox"/> QUARTZ STANDARD	
<input checked="" type="checkbox"/> FLOW DETECTOR	<input checked="" type="checkbox"/> CALIBRATION	
<input checked="" type="checkbox"/> PUMP HIGH SPEED	<input checked="" type="checkbox"/> PRINTER	
<input checked="" type="checkbox"/> INDICATOR LIGHTS		
<input checked="" type="checkbox"/> SIMULATOR SOLUTION SUPPLIER RepCo Marketing Inc.	LOT # 12002	EXP. DATE 08/29/2014
<input checked="" type="checkbox"/> SIMULATOR TEMP (34°C ± 0.2°C) 34.0 °C	SIMULATOR SN G6768	EXP. DATE 02/14/2014

CALIBRATION CHECK - (ONLY ONE STANDARD IS TO BE USED PER MAINTENANCE REPORT)

Run three tests using a standard solution. All three tests must be within ±5% of the standard value and must have a spread of .005 or less. Mark the box corresponding to the standard solution being used. (PRINTOUT ATTACHED)

- 0.100% STANDARD - MUST READ BETWEEN 0.095% AND 0.105% INCLUSIVE
- 0.080% STANDARD - MUST READ BETWEEN 0.078% AND 0.084% INCLUSIVE
- 0.040% STANDARD - MUST READ BETWEEN 0.038% AND 0.042% INCLUSIVE

TEST 1 ← .099	TEST 2 ← .101	TEST 3 ← .101
---------------	---------------	---------------

PERFORM R.E.I. TEST (PRINTOUT ATTACHED)

INDICATE THE NUMBER OF BREATH TESTS IN THE FOLLOWING RANGES SINCE THE LAST MAINTENANCE REPORT:  
(DO NOT INCLUDE SELF-ADMINISTERED TESTS)

REFUSALS 0	(0-0) 0	(.06-.09) 0	(.10-.14) 0	(.15-.19) 0	OVER .19 1
------------	---------	-------------	-------------	-------------	------------

LIST ANY NEW PARTS AND DESCRIBE ANY ALTERATION OR MODIFICATION THAT WAS MADE TO RESTORE THE INSTRUMENT TO OPERATE SATISFACTORILY AND WITHIN ESTABLISHED LIMITS (LIST OTHER SIDE IF NECESSARY).

INSPECTING OFFICER

SIGNATURE

TYPE I PERMIT NUMBER EXPIRATION DATE  
220404/02-09-2014

PRINT OUT NAME  
TROOPER P. C. LONG

TELEPHONE NUMBER  
(636) 300-2800

RETURN COMPLETED REPORT TO THE: Breath Alcohol Program, MO Department of Health and Senior Services, Southeast District Office  
2875 James Blvd.  
Poplar Bluff, MO 63001

*REPCO MARKETING INC.*

ANALYST: LINDY BRIGGS, CSE#1  
RA: FISH HAT, FISHING  
DATE: 08/26/2012

CERTIFICATE OF ANALYSIS

MANUFACTURER AND SUPPLIER: RepCo Marketing, Inc.

LOT NUMBER: 12002

EXPIRATION DATE: August 29, 2014 at 11:59 pm.

RepCo Marketing, Inc. certifies the following:

RepCo Marketing, Inc. manufactured, tested and supplied Lot Number 12002, 6% Alcohol Certified Solution for simulators. Random samples of said lot number were analyzed by an independent laboratory utilizing a gas chromatograph and found to contain .120% gms/dl +/- .003 gms/dl wt/vol ethanol (95% Confidence).

The alcohol and distilled water used in the solution were found to be free of any interfering substance.

This solution will produce a vapor alcohol value of .100 +/- 3% gms/210L Breath when heated to 34 Degrees Celsius +/- 0.2 Degrees Celsius in a simulator (95% Confidence).

The date of manufacture for this lot number is August 30, 2012.  
The expiration date for this lot number is August 29, 2014 at  
11:59 pm.

This document is a true representation of the original Certificate of Analysis.

Cecil B. Garner

Cecil B. Garner, President  
RepCo Marketing, Inc.

**BAC DataMaster**  
Evidence Ticket

Operator Signature

TPE DC Long

Witness Signature

TPE DC Long

Operator Signature

TPE DC Long

State of Missouri  
DEPARTMENT OF HEALTH

P E R M I T  
TYPE II



PAUL C LONG

Is hereby authorized to instruct and supervise operators, train instructors, inspect, calibrate, perform field repairs, and operate the following breath analyzer(s):

DATAMASTER

for the determination of the alcoholic content of blood from a sample of expired (alveolar) air, issued under the provisions of sections 577.020 through 577.041, RSMo 1986.

Date 02/09/2012

Number 220040

Expires 02/09/2014

MO 680-0771 (7-48)

*[Signature]*  
Director of State Public Health Laboratory

*Maryann T. Denney*

Director, Department of Health

Box 4 (R7-48)